



# WASHINGTON STATE PATROL SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 28, 2006 INITIALS: SEC OWNER: WSP  
 DISTRICT: 8 ADDRESS: \_\_\_\_\_  
 LOCATION: 771 Shop COUNTY: \_\_\_\_\_  
 MAKE: Haenni TYPE: port wheel weighing ACCURACY CLASS: III  
 MODEL: 101 ID. NO: 23115 CAPACITY: 20,000 lbs SECTIONS: 1  
 MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15050	+50				
Center	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.  
 Signed by: David Crona Date: 3/28/06

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed) before  
 me on March 28, 2006 by  
Michael J. Evans  
 (Signature of notary public)  
Michael J. EVANS  
 Printed Name of Notary  
 My appointment expires 12-27-07

